


StandbyU Shield Pendant

Quick Testing Guide - Monitoring Centre Plan




Prepare to Test

- Charge the device in the charging cradle.
- Let Responders know you're testing.



Call the Back2Base Monitoring Centre

- Give your name, device ID and moncode.
- Let the operator know you're going to test the device.



Activate the SOS Alarm

- Press the SOS button 3 times quickly & firmly.
- The device will vibrate 5 times.
- Speak clearly for 30 - 60 seconds.



Check back with the Back2Base Operator

- Issues during testing? Let the Operator know you'll test again in 10 minutes.
- Ask the Operator to keep the device in test/offline mode.
- If the second test isn't successful, contact the StandbyU Shield Team on 1800 069 010.